

Results:

Test Article Number	Percent BFE (%)
1	99.9
2	>99.9 ^a
3	>99.9 ^a
4	99.9
5	99.9

^a There were no detected colonies on any of the Andersen sampler plates for this test article.

Test Article Number	Delta P (mm H ₂ O/cm ²)	Delta P (Pa/cm ²)
1	6.9	67.5
2	6.7	65.6
3	6.7	65.9
4	6.7	65.7
5	6.7	65.5

The filtration efficiency percentages were calculated using the following equation:

$$\% BFE = \frac{C - T}{C} \times 100$$

C = Positive control average

T = Plate count total recovered downstream of the test article

Note: The plate count total is available upon request